

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00400 Calib. Date: 11/12/2013 Calib. No.: 00017
Certification File No.: 00401 Cert. Date: 11/12/2013 Cert. No.: 00012
Linearity File No.: 00402 Lin. Date: 11/12/2013 Lin. No.: 00012
Solution File No.: 00419 Soln. Date: 01/03/2014 Soln. No.: 00083
Sequential File No.: 00419 File Date: 01/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 04/25/2014
Solution Control Lot: 12D101 Bottle No.: 0942

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 23:41S | | |
| Control 1 EC | 0.102% | 23:41S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 23:41S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:42S | | |
| Control 2 EC | 0.101% | 23:43S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 23:43S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:44S | | |
| Control 3 EC | 0.102% | 23:45S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 23:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 23:46S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMOL

First Name: GEORGE

MI: L

Signature: _____

P.O. [Signature] #330

Badge No.: 330

Date: 01/03/2014