

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00289	Calib. Date:	12/21/2011
Certification File No.:	00290	Cert. No.:	00013
Linearity File No.:	00291	Cert. Date:	12/21/2011
Solution File No.:	00294	Lin. No.:	00008
Sequential File No.:	00294	Lin. Date:	12/21/2011
		Soln. Date:	01/20/2012
		Soln. No.:	00058
		File Date:	01/20/2012

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0216
Control Solution %:	0.100%			Expires:	03/31/2013
Solution Control Lot:	11C088			Bottle No.:	0695

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:50S		
Control 1 EC	----%	09:51S	25.2°C	SIMULATOR TEMP. ERROR
Control 1 IR	----%	09:51S	25.2°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	----%	--:--		
Control 2 EC	----%	--:--	N/A	
Control 2 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BARRETT

First Name: TRACEY

MI: L

Signature: \_\_\_\_\_

*Jay Waddie* 310

Badge No.: 310

Date: 01/20/2012

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00289	Calib. Date: 12/21/2011
Certification File No.:	00290	Calib. No.: 00013
Linearity File No.:	00291	Cert. Date: 12/21/2011
Solution File No.:	00295	Cert. No.: 00008
Sequential File No.:	00295	Lin. Date: 12/21/2011
		Lin. No.: 00008
		Soln. Date: 01/20/2012
		Soln. No.: 00059
		File Date: 01/20/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0216
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0695

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:20S		
Control 1 EC	0.104%	11:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	11:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22S		
Control 2 EC	0.103%	11:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	11:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23S		
Control 3 EC	0.103%	11:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	11:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BARRETT

First Name: TRACEY

MI: L

Signature: \_\_\_\_\_



Badge No.: 310

Date: 01/20/2012