

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0056  
Location: LONG BRANCH POLICE DEPT  
Calibration File No.: 00220 Calib. Date: 01/10/2011 Calib. No.: 00011  
Certification File No.: 00221 Cert. Date: 01/10/2011 Cert. No.: 00006  
Linearity File No.: 00222 Lin. Date: 01/10/2011 Lin. No.: 00006  
Solution File No.: 00225 Soln. Date: 02/18/2011 Soln. No.: 00045  
Sequential File No.: 00225 File Date: 02/18/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217  
Control Solution %: 0.100% Expires: 08/05/2011  
Solution Control Lot: 09H068 Bottle No.: 0237

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:19S		
Control 1 EC	0.101%	05:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:21S		
Control 2 EC	0.099%	05:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:22S		
Control 3 EC	0.099%	05:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GARRETT

First Name: BRIAN

MI: A

Signature: \_\_\_\_\_

*Brian Garrett #306*

Badge No.: 306

Date: 02/18/2011