

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00545	Calib. Date: 10/16/2015
Certification File No.:	00546	Calib. No.: 00022
Linearity File No.:	00547	Cert. Date: 10/16/2015
Solution File No.:	00572	Cert. No.: 00017
Sequential File No.:	00572	Lin. Date: 10/16/2015
		Lin. No.: 00017
		Soln. Date: 02/22/2016
		Soln. No.: 00117
		File Date: 02/22/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0217
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 1345

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:14S		
Control 1 EC	0.099%	00:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:16S		
Control 2 EC	0.098%	00:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:18S		
Control 3 EC	0.097%	00:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JOYCE

First Name: PATRICK

MI: D

Signature: _____

Handwritten Signature H 326

Badge No.: 326

Date: 02/22/2016