

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00473	Calib. No.: 00019
Certification File No.:	00474	Cert. No.: 00014
Linearity File No.:	00475	Lin. No.: 00014
Solution File No.:	00493	Soln. No.: 00100
Sequential File No.:	00493	File Date: 03/02/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0216
Control Solution %:	0.100%			Expires:	01/06/2016
Solution Control Lot:	14A126			Bottle No.:	0281

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:08S		
Control 1 EC	0.099%	01:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:09S		
Control 2 EC	0.099%	01:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:11S		
Control 3 EC	0.098%	01:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WALLS

First Name: CHRIS

MI:

Signature: _____



Badge No.: 314

Date: 03/02/2015