

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00220 Calib. Date: 01/10/2011 Calib. No.: 00011
Certification File No.: 00221 Cert. Date: 01/10/2011 Cert. No.: 00006
Linearity File No.: 00222 Lin. Date: 01/10/2011 Lin. No.: 00006
Solution File No.: 00228 Soln. Date: 03/23/2011 Soln. No.: 00046
Sequential File No.: 00228 File Date: 03/23/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217
Control Solution %: 0.100% Expires: 08/05/2011
Solution Control Lot: 09H068 Bottle No.: 0239

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:03S		
Control 1 EC	0.098%	17:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:04S		
Control 2 EC	0.097%	17:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:06S		
Control 3 EC	0.098%	17:06S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BARRETT

First Name: TRACEY

MI: L

Signature: _____



Badge No.: 310

Date: 03/23/2011