

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00629 Calib. Date: 10/21/2016 Calib. No.: 00025
Certification File No.: 00630 Cert. Date: 10/21/2016 Cert. No.: 00019
Linearity File No.: 00631 Lin. Date: 10/21/2016 Lin. No.: 00019
Solution File No.: 00664 Soln. Date: 04/03/2017 Soln. No.: 00136
Sequential File No.: 00664 File Date: 04/03/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 1326

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:38D		
Control 1 EC	0.099%	04:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:39D		
Control 2 EC	0.100%	04:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:41D		
Control 3 EC	0.099%	04:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: UMANA

First Name: HECTOR

MI: D

Signature: _____

P.O.

328

Badge No.: 328

Date: 04/03/2017