

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00350 Calib. Date: 11/30/2012 Calib. No.: 00015
Certification File No.: 00351 Cert. Date: 11/30/2012 Cert. No.: 00010
Linearity File No.: 00352 Lin. Date: 11/30/2012 Lin. No.: 00010
Solution File No.: 00362 Soln. Date: 04/14/2013 Soln. No.: 00074
Sequential File No.: 00362 File Date: 04/14/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 04/25/2014
Solution Control Lot: 12D101 Bottle No.: 0947

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:49D		
Control 1 EC	0.097%	06:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:51D		
Control 2 EC	0.098%	06:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	06:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:52D		
Control 3 EC	0.098%	06:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KENNEDY

First Name: JOSEPH

MI: V

Signature: _____



Badge No.: 324

Date: 04/14/2013