

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00289	Calib. Date:	12/21/2011
Certification File No.:	00290	Cert. Date:	12/21/2011
Linearity File No.:	00291	Lin. Date:	12/21/2011
Solution File No.:	00307	Soln. Date:	04/27/2012
Sequential File No.:	00307	File Date:	04/27/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	11C088	Expires:	03/31/2013
		Bottle No.:	0693

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:15D		
Control 1 EC	0.102%	10:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		
Control 2 EC	0.101%	10:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	10:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18D		
Control 3 EC	0.101%	10:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	10:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WIDDIS

First Name: TRACEY

MI: L

Signature: \_\_\_\_\_



Badge No.: 310

Date: 04/27/2012