

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00499	Calib. Date:	04/08/2015
Certification File No.:	00500	Cert. Date:	04/08/2015
Linearity File No.:	00501	Lin. Date:	04/08/2015
Solution File No.:	00517	Soln. Date:	06/08/2015
Sequential File No.:	00517	File Date:	06/08/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0216
Control Solution %:	0.100%			Expires:	01/06/2016
Solution Control Lot:	14A126			Bottle No.:	0280

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:31D		
Control 1 EC	0.101%	23:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:33D		
Control 2 EC	0.100%	23:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:34D		
Control 3 EC	0.099%	23:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: UMANA

First Name: HECTOR

MI: D

Signature: \_\_\_\_\_

Badge No.: 328

Date: 06/08/2015