

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00593	Calib. Date: 04/22/2016
Certification File No.:	00594	Calib. No.: 00024
Linearity File No.:	00595	Cert. Date: 04/22/2016
Solution File No.:	00607	Cert. No.: 00018
Sequential File No.:	00607	Lin. Date: 04/22/2016
		Lin. No.: 00018
		Soln. Date: 07/16/2016
		Soln. No.: 00123
		File Date: 07/16/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0216
Solution Control Lot:	15H140	Expires: 08/04/2017
		Bottle No.: 1328

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:59D		
Control 1 EC	0.099%	06:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:00D		
Control 2 EC	0.099%	06:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:02D		
Control 3 EC	0.099%	06:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WIDDIS

First Name: MATTHEW

MI: J

Signature: _____



Badge No.: 325

Date: 07/16/2016