

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00499 Calib. Date: 04/08/2015 Calib. No.: 00020
Certification File No.: 00500 Cert. Date: 04/08/2015 Cert. No.: 00015
Linearity File No.: 00501 Lin. Date: 04/08/2015 Lin. No.: 00015
Solution File No.: 00522 Soln. Date: 07/17/2015 Soln. No.: 00105
Sequential File No.: 00522 File Date: 07/17/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 01/06/2016
Solution Control Lot: 14A126 Bottle No.: 0284

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:59D		
Control 1 EC	0.096%	21:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:00D		
Control 2 EC	0.096%	21:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02D		
Control 3 EC	0.096%	21:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: _____



Badge No.: 336

Date: 07/17/2015