

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0056  
Location: LONG BRANCH POLICE DEPT  
Calibration File No.: 00256 Calib. Date: 07/05/2011 Calib. No.: 00012  
Certification File No.: 00257 Cert. Date: 07/05/2011 Cert. No.: 00007  
Linearity File No.: 00258 Lin. Date: 07/05/2011 Lin. No.: 00007  
Solution File No.: 00263 Soln. Date: 08/02/2011 Soln. No.: 00052  
Sequential File No.: 00263 File Date: 08/02/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216  
Control Solution %: 0.100% Expires: 03/31/2013  
Solution Control Lot: 11C088 Bottle No.: 1376

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:35D		
Control 1 EC	0.101%	16:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:36D		
Control 2 EC	0.100%	16:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:37D		
Control 3 EC	0.100%	16:38D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BARRETT

First Name: TRACEY

MI: L

Signature: \_\_\_\_\_



Badge No.: 310

Date: 08/02/2011