

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00499	Calib. Date: 04/08/2015
Certification File No.:	00500	Calib. No.: 00020
Linearity File No.:	00501	Cert. Date: 04/08/2015
Solution File No.:	00529	Cert. No.: 00015
Sequential File No.:	00529	Lin. Date: 04/08/2015
		Lin. No.: 00015
		Soln. Date: 08/17/2015
		Soln. No.: 00107
		File Date: 08/17/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXN S3-0216
Control Solution %:	0.100%		Expires: 01/06/2016
Solution Control Lot:	14A126		Bottle No.: 0282

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:55D		
Control 1 EC	0.100%	15:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:56D		
Control 2 EC	0.100%	15:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:58D		
Control 3 EC	0.099%	15:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: VALDES

First Name: HEATHER

MI: -

Signature: _____



Badge No.: 290

Date: 08/17/2015