

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00364	Calib. Date:	05/21/2013
Certification File No.:	00365	Cert. Date:	05/21/2013
Linearity File No.:	00366	Lin. Date:	05/21/2013
Solution File No.:	00387	Soln. Date:	09/03/2013
Sequential File No.:	00387	File Date:	09/03/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	12D101	Expires:	04/25/2014
		Bottle No.:	0936

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:22D		
Control 1 EC	0.099%	16:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:24D		
Control 2 EC	0.098%	16:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:25D		
Control 3 EC	0.098%	16:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: VALDES

First Name: HEATHER

MI: -

Signature: _____



Badge No.: 290

Date: 09/03/2013