

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00682	Calib. Date:	09/14/2017
Certification File No.:	00683	Cert. Date:	09/14/2017
Linearity File No.:	00684	Lin. Date:	09/14/2017
Solution File No.:	00694	Soln. Date:	10/15/2017
Sequential File No.:	00694	File Date:	10/15/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	17200	Expires:	06/28/2019
		Bottle No.:	1440

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:20D		
Control 1 EC	0.100%	06:20D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21D		
Control 2 EC	0.101%	06:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	06:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		
Control 3 EC	0.100%	06:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	06:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WALLS

First Name: CHRIS

MI:

Signature: _____

Badge No.: 314

Date: 10/15/2017