

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00256	Calib. Date: 07/05/2011
Certification File No.:	00257	Calib. No.: 00012
Linearity File No.:	00258	Cert. Date: 07/05/2011
Solution File No.:	00282	Cert. No.: 00007
Sequential File No.:	00282	Lin. Date: 07/05/2011
		Lin. No.: 00007
		Soln. Date: 11/08/2011
		Soln. No.: 00055
		File Date: 11/08/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0216
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0698

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:22S		
Control 1 EC	0.101%	17:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:23S		
Control 2 EC	0.101%	17:24S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	17:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:25S		
Control 3 EC	0.101%	17:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BARRETT

First Name: TRACEY

MI: L

Signature: _____



Badge No.: 310

Date: 11/08/2011