

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-------------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0056 |
| Location: | LONG BRANCH POLICE DEPT | |
| Calibration File No.: | 00682 | Calib. Date: 09/14/2017 |
| Certification File No.: | 00683 | Calib. No.: 00027 |
| Linearity File No.: | 00684 | Cert. Date: 09/14/2017 |
| Solution File No.: | 00697 | Cert. No.: 00021 |
| Sequential File No.: | 00697 | Lin. Date: 09/14/2017 |
| | | Lin. No.: 00021 |
| | | Soln. Date: 11/10/2017 |
| | | Soln. No.: 00142 |
| | | File Date: 11/10/2017 |

| | | |
|-----------------------------|------------------|--------------------------|
| Calibrating Unit: WET | Model No.: CU-34 | Serial No.: DDXN S3-0217 |
| Control Solution %: 0.100% | | Expires: 06/28/2019 |
| Solution Control Lot: 17200 | | Bottle No.: 1439 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|-----------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 18:41S | | |
| Control 1 EC | -.---% | 18:42S | 34.4°C | SIMULATOR TEMP. ERROR |
| Control 1 IR | -.---% | 18:42S | 34.4°C | SIMULATOR TEMP. ERROR |
| Ambient Air Blank | -.---% | --:-- | | |
| Control 2 EC | -.---% | --:-- | N/A | |
| Control 2 IR | -.---% | --:-- | N/A | |
| Ambient Air Blank | -.---% | --:-- | | |
| Control 3 EC | -.---% | --:-- | N/A | |
| Control 3 IR | -.---% | --:-- | N/A | |
| Ambient Air Blank | -.---% | --:-- | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMOL

First Name: GEORGE

MI: L

Signature: _____



Badge No.: 330

Date: 11/10/2017