

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00256	Calib. Date:	07/05/2011
Certification File No.:	00257	Cert. Date:	07/05/2011
Linearity File No.:	00258	Lin. Date:	07/05/2011
Solution File No.:	00287	Soln. Date:	12/09/2011
Sequential File No.:	00287	File Date:	12/09/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	11C088	Expires:	03/31/2013
		Bottle No.:	0699

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:19S		
Control 1 EC	0.102%	16:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:21S		
Control 2 EC	0.100%	16:21S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:22S		
Control 3 EC	0.101%	16:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	16:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: BARRETT

First Name: TRACEY

MI: L

Signature: _____



Badge No.: 310

Date: 12/09/2011