

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
 Location: LONG BRANCH POLICE DEPT
 Calibration File No.: 00400 Calib. Date: 11/12/2013 Calib. No.: 00017
 Certification File No.: 00401 Cert. Date: 11/12/2013 Cert. No.: 00012
 Linearity File No.: 00402 Lin. Date: 11/12/2013 Lin. No.: 00012
 Solution File No.: 00410 Soln. Date: 12/09/2013 Soln. No.: 00082
 Sequential File No.: 00410 File Date: 12/09/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
 Control Solution %: 0.100% Expires: 04/25/2014
 Solution Control Lot: 12D101 Bottle No.: 0932

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:57S		
Control 1 EC	0.098%	12:58S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:59S		
Control 2 EC	0.098%	12:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:00S		
Control 3 EC	0.098%	13:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXA 92-114 MD

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SGT. [Signature] #5925

Badge No.: 5925

Date: 12/09/2013