

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00350 Calib. Date: 11/30/2012 Calib. No.: 00015
Certification File No.: 00351 Cert. Date: 11/30/2012 Cert. No.: 00010
Linearity File No.: 00352 Lin. Date: 11/30/2012 Lin. No.: 00010
Solution File No.: 00356 Soln. Date: 01/20/2013 Soln. No.: 00072
Sequential File No.: 00356 File Date: 01/20/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 04/25/2014
Solution Control Lot: 12D101 Bottle No.: 0939

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:30S		
Control 1 EC	0.099%	21:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	21:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:32S		
Control 2 EC	0.097%	21:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:34S		
Control 3 EC	0.097%	21:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WALLS

First Name: CHRIS

MI:

Signature: _____

Chris Walls 314

Badge No.: 314

Date: 01/20/2013