

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00682 Calib. Date: 09/14/2017 Calib. No.: 00027
Certification File No.: 00683 Cert. Date: 09/14/2017 Cert. No.: 00021
Linearity File No.: 00684 Lin. Date: 09/14/2017 Lin. No.: 00021
Solution File No.: 00710 Soln. Date: 01/26/2018 Soln. No.: 00147
Sequential File No.: 00710 File Date: 01/26/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 06/28/2019
Solution Control Lot: 17200 Bottle No.: 1452

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:09S		
Control 1 EC	0.100%	18:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:11S		
Control 2 EC	0.100%	18:11S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:12S		
Control 3 EC	0.100%	18:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: VALDES

First Name: HEATHER

MI: -

Signature: _____



Badge No.: 290

Date: 01/26/2018