

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00545 **Calib. Date:** 10/16/2015 **Calib. No.:** 00022
Certification File No.: 00546 **Cert. Date:** 10/16/2015 **Cert. No.:** 00017
Linearity File No.: 00547 **Lin. Date:** 10/16/2015 **Lin. No.:** 00017
Solution File No.: 00565 **Soln. Date:** 01/30/2016 **Soln. No.:** 00113
Sequential File No.: 00565 **File Date:** 01/30/2016

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXN S3-0217
Control Solution %: 0.100% **Expires:** 08/14/2017
Solution Control Lot: 15H140 **Bottle No.:** 1335

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:14S		
Control 1 EC	0.099%	03:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:16S		
Control 2 EC	0.098%	03:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:18S		
Control 3 EC	0.099%	03:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JOYCE

First Name: PATRICK

MI: D

Signature: _____



Badge No.: 326

Date: 01/30/2016