

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00289 Calib. Date: 12/21/2011 Calib. No.: 00013
Certification File No.: 00290 Cert. Date: 12/21/2011 Cert. No.: 00008
Linearity File No.: 00291 Lin. Date: 12/21/2011 Lin. No.: 00008
Solution File No.: 00305 Soln. Date: 03/30/2012 Soln. No.: 00061
Sequential File No.: 00305 File Date: 03/30/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0696

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:38S		
Control 1 EC	0.101%	10:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40S		
Control 2 EC	0.101%	10:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41S		
Control 3 EC	0.101%	10:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WIDDIS

First Name: TRACEY

MI: L

Signature: _____



Badge No.: 310

Date: 03/30/2012