

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00593	Calib. Date:	04/22/2016
Certification File No.:	00594	Cert. Date:	04/22/2016
Linearity File No.:	00595	Lin. Date:	04/22/2016
Solution File No.:	00600	Soln. Date:	05/06/2016
Sequential File No.:	00600	File Date:	05/06/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	1329

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:07D		
Control 1 EC	0.100%	03:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:09D		
Control 2 EC	0.100%	03:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:10D		
Control 3 EC	0.100%	03:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: JOYCE

First Name: PATRICK

MI: D

Signature: _____



Badge No.: 326

Date: 05/06/2016