

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00220	Calib. Date:	01/10/2011
Certification File No.:	00221	Cert. Date:	01/10/2011
Linearity File No.:	00222	Lin. Date:	01/10/2011
Solution File No.:	00242	Soln. Date:	05/12/2011
Sequential File No.:	00242	File Date:	05/12/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	09H068	Expires:	08/05/2011
		Bottle No.:	0247

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:54D		
Control 1 EC	0.098%	00:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	00:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:55D		
Control 2 EC	0.097%	00:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	00:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:57D		
Control 3 EC	0.097%	00:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GARRETT

First Name: BRIAN

MI: A

Signature: \_\_\_\_\_

Badge No.: 306

Date: 05/12/2011