

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00364	Calib. Date:	05/21/2013
Certification File No.:	00365	Cert. No.:	00016
Linearity File No.:	00366	Cert. Date:	05/21/2013
Solution File No.:	00370	Lin. No.:	00011
Sequential File No.:	00370	Soln. Date:	06/27/2013
		Soln. No.:	00076
		File Date:	06/27/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	12D101	Expires:	04/25/2014
		Bottle No.:	0945

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:51D		
Control 1 EC	0.100%	17:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52D		
Control 2 EC	0.099%	17:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:54D		
Control 3 EC	0.099%	17:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SAMOL

First Name: GEORGE

MI: L

Signature: \_\_\_\_\_

*PO. [Signature] #330*

Badge No.: 330

Date: 06/27/2013

*#330*