

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00311	Calib. Date: 06/06/2012
Certification File No.:	00312	Calib. No.: 00014
Linearity File No.:	00313	Cert. Date: 06/06/2012
Solution File No.:	00319	Cert. No.: 00009
Sequential File No.:	00319	Lin. Date: 06/06/2012
		Lin. No.: 00009
		Soln. Date: 07/08/2012
		Soln. No.: 00065
		File Date: 07/08/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0217
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0688

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:24D		
Control 1 EC	0.099%	16:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:25D		
Control 2 EC	0.099%	16:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27D		
Control 3 EC	0.100%	16:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: VALDES

First Name: HEATHER

MI: -

Signature: \_\_\_\_\_



Badge No.: 290

Date: 07/08/2012