

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00311	Calib. Date:	06/06/2012
Certification File No.:	00312	Cert. Date:	06/06/2012
Linearity File No.:	00313	Lin. Date:	06/06/2012
Solution File No.:	00323	Soln. Date:	07/30/2012
Sequential File No.:	00323	File Date:	07/30/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	11C088	Expires:	03/31/2013
		Bottle No.:	0686

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:49D		
Control 1 EC	0.098%	11:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50D		
Control 2 EC	0.099%	11:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		
Control 3 EC	0.098%	11:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WIDDIS

First Name: TRACEY

MI: L

Signature: _____

Badge No.: 310

Date: 07/30/2012