

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00311 Calib. Date: 06/06/2012 Calib. No.: 00014
Certification File No.: 00312 Cert. Date: 06/06/2012 Cert. No.: 00009
Linearity File No.: 00313 Lin. Date: 06/06/2012 Lin. No.: 00009
Solution File No.: 00331 Soln. Date: 08/29/2012 Soln. No.: 00067
Sequential File No.: 00331 File Date: 08/29/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0691

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:03D		
Control 1 EC	0.100%	15:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:05D		
Control 2 EC	0.099%	15:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06D		
Control 3 EC	0.098%	15:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WIDDIS

First Name: TRACEY

MI: L

Signature: PO J. Widdis 310

Badge No.: 310

Date: 08/29/2012