

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARXA-0056 |
| Location: | LONG BRANCH POLICE DEPT | | |
| Calibration File No.: | 00593 | Calib. Date: | 04/22/2016 |
| Certification File No.: | 00594 | Cert. Date: | 04/22/2016 |
| Linearity File No.: | 00595 | Lin. Date: | 04/22/2016 |
| Solution File No.: | 00628 | Soln. Date: | 10/09/2016 |
| Sequential File No.: | 00628 | File Date: | 10/09/2016 |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDXN S3-0216 |
| Solution Control Lot: | 15H140 | Expires: | 08/04/2017 |
| | | Bottle No.: | 1343 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 06:15D | | |
| Control 1 EC | 0.100% | 06:16D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 06:16D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:16D | | |
| Control 2 EC | 0.100% | 06:17D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 06:17D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:18D | | |
| Control 3 EC | 0.099% | 06:19D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 06:19D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 06:19D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: JOYCE

First Name: PATRICK

MI: D

Signature: _____



Badge No.: 326

Date: 10/09/2016