

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00545	Calib. Date:	10/16/2015
Certification File No.:	00546	Cert. No.:	00022
Linearity File No.:	00547	Cert. Date:	10/16/2015
Solution File No.:	00554	Lin. No.:	00017
Sequential File No.:	00554	Lin. Date:	10/16/2015
		Soln. No.:	00111
		Soln. Date:	11/19/2015
		File Date:	11/19/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0217
Control Solution %:	0.100%			Expires:	01/06/2016
Solution Control Lot:	14A126			Bottle No.:	1324

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:33S		
Control 1 EC	0.100%	03:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:35S		
Control 2 EC	0.099%	03:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:36S		
Control 3 EC	0.098%	03:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: UMANA

First Name: HECTOR

MI: D

Signature: _____

Badge No.: 328

Date: 11/19/2015